Sensitive NMR Sensors Detect Antibodies to Influenza

Citation

Koh, Isaac, Rui Hong, Ralph Weissleder, and Lee Josephson. 2008. "Sensitive NMR Sensors Detect Antibodies to Influenza." *Angewandte Chemie International Edition* 47 (22): 4119–21. <https://doi.org/10.1002/anie.200800069>.

Permanent link

<http://nrs.harvard.edu/urn-3:HUL.InstRepos:41384406>

Terms of Use

This article was downloaded from Harvard University's DASH repository, and is made available under the terms and conditions applicable to Other Posted Material, as set forth at <http://nrs.harvard.edu/urn-3:HUL.InstRepos:dash.current.terms-of-use#LAA>

Share Your Story

The Harvard community has made this article openly available.
Please share how this access benefits you. [Submit a story](#).

[Accessibility](#)

Published in final edited form as:

Angew Chem Int Ed Engl. 2008 ; 47(22): 4119–4121. doi:10.1002/anie.200800069.

Sensitive NMR Sensors Detect Antibodies To Influenza

Isaac Koh, Rui Hong, Ralph Weissleder, and Lee Josephson*

Center for Molecular Imaging Research Massachusetts General Hospital and Harvard Medical School Building 149, 13th Street, Boston, MA 02129 (USA)

Keywords

antibodies; magnetic properties; peptides; biosensors; virus

Point of care (POC) sensors for the determination of immune status are an important part of the preparation for possible pandemics such as that created by avian influenza.[1,2] NMR based magnetic relaxation switches (MRSw's) are attractive for this application because they are indifferent to light and involve no immobilization of materials on vessel walls.[3] MR relaxometers are practical as POC sensors since their requirements for magnetic field strength, volume and homogeneity are minimal.[4-6]

To date, MRSw's have used magnetic nanoparticles (NPs) that react with molecular targets to aggregate and decrease T_2 . Although MRSw's detect highly multivalent viruses or bacteria with high sensitivity,[7,8] their sensitivity for proteins is far lower.[9,10] Our goal was to improve the sensitivity of MRSw sensors for divalent antibodies, so the immune status of birds or humans might eventually be determined using a POC MR relaxometer. A monoclonal antibody recognizing the Tag peptide from a hemagglutinin of a human influenza virus was used as a target to assess strategies for improving MRSw sensitivity.

An obvious strategy for improving MRSw sensor sensitivity was to employ the equivalence principle of antibody-antigen reactions involved in precipitin formation,[11-13] reducing the concentration of Tag peptide-magnetic particles (a synthetic multivalent antigen), so that lower concentrations of anti-Tag antibodies might achieve aggregation. The Tag peptide from the influenza hemagglutinin was therefore conjugated to nanoparticles (NPs) and far larger micron-sized magnetic particles (MPs), to yield Tag-NPs and Tag-MPs. As indicated in Table 1, a Tag-MP contained 350,000 times more iron than a Tag-NP and had a correspondingly higher magnetic moment per particle. With a typical initial T_2 for MRSw assays (100 msec), the concentration of particles decreased from 2.8×10^{-9} M with Tag-NPs to 5.1×10^{-15} M using Tag-MPs.

Attempts to use Tag-MPs for MRSw assays indicated Tag-MPs differed from earlier NPs in two key respects. First, as shown in Figure 1, Tag-MP's underwent a reversible increase in T_2 in the 0.47 T relaxometer magnet. Micrographs indicated the T_2 increase was associated with MP aggregation resulting from magnetic attractions between Tag-MP's (see Figure 2 of

*corresponding author Fax: (+1)617-726-5708 E-mail: josephso@helix.mgh.harvard.edu.

Justification

New sensors for the detection of antibodies to the structural features of viral proteins are an important part of the preparation for possible viral pandemics. Here we show that micron-sized particles that are 350,000 times more magnetic (per particle) than earlier magnetic nanoparticles provide a high sensitivity, NMR based sensor for detecting antibodies to influenza virus. A homogeneous magnetic field was employed to enhance antibody-based crosslinking between particles, providing a novel method of increasing the sensitivity of a homogeneous (non-separation based) immunoassay.

reference [14]), with the randomizing effects of thermal energy causing dispersion when the field was removed. Second, Tag-MP aggregation resulted in a T_2 increase while Tag-NP aggregation resulted in a T_2 decrease,[3,15] observations consistent with the outer sphere diffusion theory used to describe the effects of magnetic particles on T_2 .[16-19] This theory employs two parameters, τ_d , the diffusion time for water, and $\Delta\omega$, the difference in angular frequency between the magnetic field experienced by a proton at the particle (or aggregate) surface versus that in the bulk. Outer sphere diffusion theory predicts T_2 will decrease as Tag-NPs aggregate, since the motional averaging condition is fulfilled with both dispersed and aggregated materials ($\Delta\omega * \tau_d < 1$). With Tag-MP aggregation, the resulting magnetic field inhomogeneities become so few and infrequent water molecules must traverse long distances to encounter them. Here the effects on T_2 become diffusion limited ($\Delta\omega * \tau_d > 1$). The effects of NP and MP aggregation on T_2 have been discussed.[14]

Three techniques were then explored separately and in combination to increase the sensitivity of MRSw assays, see Figure 2. First, a *decrease in particle concentration* was achieved by replacing the magnetic nanoparticle (NP) with a micron-sized particle (MP), see Figure 2a. The second technique, magnetic field enhanced, molecular target aggregation, exploited the reversible magnetic field induced MP aggregation in the absence of a target ligand shown in Figure 1. We hypothesized that the magnetic field aggregation of Tag-MPs might enhance anti-Tag mediated aggregation (Figure 2b). Therefore solutions of Tag-MPs were exposed to anti-Tag in the relaxometer magnet, resulting in both magnetic field induced and anti-Tag antibody mediated Tag-MP aggregation. Samples were then removed from the magnetic field, to allow Tag-MPs to disaggregate which occurred only in the absence of particle crosslinking by anti-Tag. T_2 was determined in less than 30 seconds by placing samples in the relaxometer so that magnetic field induced aggregation during T_2 measurement was minimal. Finally, we employed *valency enhancement*, where the valency of the monoclonal anti-Tag was increased above the normal two antigen-combining sites per IgG by adding a sheep F(ab')₂ antibody to the Fc fragment of the mouse anti-Tag monoclonal antibody (Figure 2c).

We first measured T_2 as a function of anti-Tag concentration using Tag-NPs and Tag-MPs (decreasing particle concentration but without magnetic field or valency enhancement) as shown in Figure 3a. Data were fit to a 4-parameter equation, to obtain the EC50 (midpoint), the Hill coefficient (curve slope), and computer generated maximum and minimum values of T_2 . Table 2 provides EC50's, Hill coefficients, and ΔT_2 's for the curves shown in Figure 3a. ΔT_2 , the T_2 for the fully dispersed state minus the T_2 for the fully aggregated state, is positive for anti-Tag induced Tag-NP aggregation but negative for anti-Tag induced Tag-MP aggregation. % ΔT_2 is the difference in T_2 in the presence and absence of anti-Tag divided by ΔT_2 and expressed in percent.

Using Tag-NPs, increasing concentrations of anti-Tag gave an EC50 of 49 nM, versus an EC50 of 0.60 nM for Tag-MPs. Further reductions in the EC50's of the MP based MRSw assays were obtained by the use of magnetic field and valency enhancement as shown in Figure 3b. Through a combination of all three techniques, the EC50's for anti-Tag were progressively decreased from a starting value of 49 nM with Tag-NPs to 0.0020 nM (2.0 pM), obtained with decreased particle concentration, magnetic field aggregation, and valency enhancement (Figure 3c).

To obtain a detection limit from these results, we determined a "projected sensitivity concentration" or PSC. Five msec was added to (or subtracted from) the value of T_2 in the absence of anti-Tag and the concentration at this T_2 determined from curve parameters (Table 2). The PSC for anti-Tag with Tag-NP sensor, the starting point of our investigation, was 26 nM. By using our three techniques for enhancing sensitivity, the PSC was reduced to 0.00014 nM or 186,000 fold. The PSC is discussed further in Supporting Information. This detection

sensitivity is comparable with those of most advanced nanostructure-based antibody assays. [20]

We obtained a high sensitivity MRSw sensor for influenza antibody by using MPs and exploiting a previously unrecognized feature of their response to a magnetic field. Application of a homogeneous magnetic field enhanced antibody-based crosslinking between particles, providing a novel method of increasing the sensitivity of a homogeneous, particle aggregation based immunoassay. Together with high throughput methods of peptide and peptide-MP conjugate synthesis,[21] MRSw sensors might be used, both to analyze the immune response to mutating viruses in laboratory settings, and for an MRSw based POC antibody sensor.

Acknowledgements

This work was supported by R01EB0004626 and U54CA119349.

References

1. Lu PS. Science 2006;312:337. [PubMed: 16627701]
2. Butler D. Nature 2006;439:248. [PubMed: 16421529]
3. Perez JM, Josephson L, O'Loughlin T, Hogemann D, Weissleder R. Nat. Biotechnol 2002;20:816. [PubMed: 12134166]
4. Manz B, Coy A, Dykstra R, Eccles CD, Hunter MW, Parkinson BJ, Callaghan PT. J. Magn. Reson 2006;183:25. [PubMed: 16891134]
5. Blumich B, Blumler P, Eidmann G, Guthausen A, Haken R, Schmitz U, Saito K, Zimmer G. Magn. Reson. Imaging 1998;16:479. [PubMed: 9803893]
6. Sillerud LO, McDowell AF, Adolphi NL, Serda RE, Adams DP, Vasile MJ, Alam TM. J. Magn. Reson 2006;181:181. [PubMed: 16698297]
7. Perez JM, Simeone FJ, Saeki Y, Josephson L, Weissleder R. J. Am. Chem. Soc 2003;125:10192. [PubMed: 12926940]
8. Kaittanis C, Naser SA, Perez JM. Nano Lett 2007;7:380. [PubMed: 17298004]
9. Kim GY, Josephson L, Langer R, Cima MJ. Bioconjugate Chem 2007;18:2024.
10. Perez JM, O'Loughlin T, Simeone FJ, Weissleder R, Josephson L. J. Am. Chem. Soc 2002;124:2856. [PubMed: 11902860]
11. DeLisi C. J. Theor. Biol 1974;45:555. [PubMed: 4135952]
12. Chak KC, Hart H. Bull. Math. Biol 1980;42:37. [PubMed: 7357120]
13. van Oss CJ. J. Immunoassay 2000;21:143. [PubMed: 10929885]
14. Hong R, Cima MJ, Weissleder R, Josephson L. Mag. Reson. Med. 2008in press
15. Josephson L, Perez JM, Weissleder R. Angew. Chem. Int. Ed 2001;40:3204.
16. Gillis P, Koenig SH. Magn. Reson. Med 1987;5:323. [PubMed: 2824967]
17. Brooks RA. Magn. Reson. Med 2002;47:388. [PubMed: 11810684]
18. Yung KT. Magn. Reson. Imaging 2003;21:451. [PubMed: 12878254]
19. Roch A, Gossuin Y, Muller RN, Gillis P. J. Magn. Magn. Mater 2005;293:532.
20. Rosi NL, Mirkin CA. Chem. Rev 2005;105:1547. [PubMed: 15826019]
21. Weissleder R, Kelly K, Sun EY, Shtatland T, Josephson L. Nat. Biotechnol 2005;23:1418. [PubMed: 16244656]
22. Reynolds F, O'Loughlin T, Weissleder R, Josephson L. Anal. Chem 2005;77:814. [PubMed: 15679348]

Figure 1. Magnetic particles respond to a homogeneous magnetic field

T_2 of a MP solution (●), but not a NP solution (□), increases when in a 0.47 T field (gray). White, no field.

a) Decreased Particle Concentration

b) Magnetic Field Enhanced Target Aggregation

c) Valency Enhanced Aggregation

Figure 2. Methods of increasing MRSw sensitivity

a) Decrease in particle concentration. Anti-Tag forms bridges between larger MPs. b) Magnetic field enhanced target aggregation. A magnetic field (+H) produces magnetic attractions between MPs that result in aggregation. In the presence of anti-Tag (top) the aggregate is maintained when the field is removed (−H) and a high T_2 results. c) Valency enhanced aggregation. The addition of an anti-Fc antibody permits anti-Tag to bind more than two MPs simultaneously.

Figure 3. Techniques for increasing the response to anti-Tag antibody

a) The initial reference assay with Tag-NPs is shown with a dotted line and (●). To obtain a decrease in particle concentration, Tag-MPs (○) were used. The Tag-MP was then used with magnetic field aggregation (□), with valency enhancement (Δ), and with both magnetic aggregation and valency enhancement (☆). b) EC_{50} 's obtained by decreasing particle concentration, magnetic field aggregation (MA), and valency enhancement (VE). Additional curve parameters are provided in Table 2.

Table 1

Properties of magnetic particles (MPs) and nanoparticles (NPs).

Particle	MP	NP
Size (nm)	1000	30
Settling	<5%	None
Peptides per particle	3.0×10^5	20-30
R2 (sec ⁻¹ /mM Fe)	43	50
M (emu/g Fe)	105	86.6
Fe atoms per particle	2.8×10^9	8000 ^[a]
Particle Conc. @ T2 = 100 msec, (M)	5.1×10^{-15}	2.8×10^{-9}

^[a]From[22].

Table 2

Sensitivity of NP and MP Based MRS_w Sensors.

	Magnetic Aggregation	Valency Enhancement	EC ₅₀ (nM)	Hill slope	ΔT_2 (msec)	PSC (nM)
Tag-NP	N	N	49	4.7	+110	26
Tag-MP	N	N	0.60	1.4	-20	0.41
	Y	N	0.55	2.1	-140	0.060
	N	Y	0.017	0.55	-64	0.00020
	Y	Y	0.0020	1.5	-250	0.00014